

SEP 1 8 2009

510(k) Summary of Safety and Effectiveness
FxDEVICES POGO Screw
July 20, 2009

1. **Sponsor Name**
FxDevices
One South Ocean Blvd., Suite 324
Boca Raton, FL 33432

2. **Device Name**
POGO Screw
Panel Orthopaedic
Classification Name Smooth or Threaded Metallic Bone Fixation
Fastener
CFR Number Class II (per 21 CFR 888.3040)
Product Code HWC

3. **Identification of Predicate or Legally Marketed Device**
The POGO Screw is substantially equivalent to the POGO Screw cleared under K080649.

4. **Device Description**
The POGO Screws are comprised of various size cannulated screws for the fixation of bone fractures. POGO screws are made of 316LVM Stainless Steel conforming to ASTM F138. The screws are available in various sizes from 55mm to 130mm in length. They are provided sterile and also non sterile to be sterilized by the user prior to use.

5. **Intended Use**
The POGO Screw is indicated for use in the general management of fractures and reconstructive surgery.

6. **Comparison of Technological Characteristics**
The POGO Screw and the predicate device accomplish the same function of providing compression fixation between a base bone and a bone fragment. Both devices accommodate a range of total lengths within each product design.

7. **Performance Testing**
Bench testing was conducted to support equivalency

8. **Statement of Equivalency**

The POGO Screw is substantially equivalent in design, materials, construction and intended use as those of the predicate. Since the POGO Screw is the same in intended use and technological characteristics as the predicate devices, the POGO Screw does not raise any new safety and efficacy concerns when compared to these similar legally marketed devices.

The risk analysis and test results demonstrate that the POGO Screw is substantially equivalent to the predicate device and is capable of safely and effectively performing the stated intended use.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center - WO66-G609
Silver Spring, MD 20993-0002

SEP 18 2009

FxDEVICES
c/o Mr. Rich Lipschutz
President, FxDEVICES
One South Ocean Boulevard, Suite 324
Boca Raton, Florida 33432

Re: K092189
Trade/Device Name: POGO Screw
Regulation Number: 21 CFR 888.3040
Regulation Name: Smooth or threaded metallic bone fixation fastener
Regulatory Class: Class II
Product Code: HWC
Dated: August 19, 2009
Received: August 21, 2009

Dear Mr. Lipschutz:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set


Page 2 - Mr. Rich Lipschutz

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Mark N. Melkerson
Director Division of Surgical, Orthopedic
and Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K092189

Device Name: POGO Screw

Indications For Use:

The POGO Screw is indicated for use in the general management of fractures and reconstructive surgery.

Prescription Use X
(Per 21 CFR 801.109)

OR

Over-The-Counter Use _____

(PLEASE DO NOT WRITE BELOW THIS LINE. CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Jmitu J. for MYM
(Division Sign-Off)
Division of Surgical, Orthopedic,
and Restorative Devices

510(k) Number K092189